

DVT TO F

TORQUE METER



MEASURES THE TORQUE STRENGTH OF PRODUCTS

DVT TO F

TECHNICAL SPECIFICATIONS

USED IN

- Torque meter calibrators.

DVT TO S

- Bottle manufacturers and users.
- Cover producers and users.

DVT TO S O

- Bottle manufacturers and users.
- Cover producers and users.
- Baby bottle, pacifier manufacturers.

DVT TO F

- Zipper manufacturers.

RELEVANT STANDARDS

- TS 7771 EN ISO 8113

| MODELS | DVT TO S | DVT TO S O | DVT TO F |
|-----------------------|----------|------------|----------|
| DIMENSIONS (WxDxH) cm | 38x26x39 | 66x21x76 | 37x23x41 |
| WEIGHT kg | 12 | 25 | 10.5 |
| FORCE N.m | 1-3 | 1-3 | 3 |

TECHNICAL INFORMATION

- Electronic measuring system.
- Digital screen.
- Produced in various capacities.
- Max. Torque is kept in PC memory.
- Different models for various types of materials' torque strength experiments.
- Operating Voltage:220 VAC- 50 Hz.

DVT TO S

- Bottle diameter range : 3-20 cm.
- Two different models, - 1 N.m or 3 N.m according to cap opening force.
- Produced for, measurement of opening force of bottle and can.

DVT TO S O

- Measures opening strength of the cap by rotating the sample.

DVT TO F

- Can be used for zipper pull tab test.
- Angle of rotation indicator.

DVT TO S



DVT TO S O



Devotrans reserves the right, to make changes to the brochure. Photographed models may not be the latest ones.